Cleaning & Sterilisation Steps



1 Add detergent to warm water, then insert the tube attached to the instrument and flush thoroughly.

2 Reflush the instrument with clean water.

3 Remove the tube and insert a small quantity of vegetable oil in the inlet fitting.

4 Fill a deep pot with clean water. Attach tube to instrument and pump handle to draw water in to fill the cylinder.

5 Suspend the whole instrument so it does not touch the base of the pot, add the needles and boil for 10 - 20 minutes. ***Failing to suspend the instrument may result in damage to the instrument.***

6 Remove instrument from the pot, then wrap a clean cloth around the handle and pump dry.

prodigy instruments

Thank you for your purchase of this product, as it continues a family tradition since 1931. From my grandfather Norman Phillips, to my father John Phillips and now to me, Matthew Phillips...the tradition lives on.

Manufactured by: MSAN GLOBAL PTY LIMITED, ABN: 85 711 822 698, P.O. Box 120, Mount Kuring-Gai, NSW, 2080, Australia.
P: +61 403 454 354 E: matthew@prodigyinstruments.com.au